

REGISTRATION OF A PRESSURE FITTING DESIGN

28-May-19

ABSA
9410 20th Avenue NW
Edmonton, AB
T6N 0A4

Attention: Cynthia Formaniuk

File Number: 11132 [0 F]

Re: **Manufacturer:** Sierra Instruments, Inc.
Item: Flow Meters and Controllers - Series 100, 800, 600, 700, & 200.
Catalog or Drawing: Scope of Registration AB-257

TSASK Codes and Standards Compliance has registered the design listed above in accordance with The Boiler and Pressure Vessel Act and Regulations and CSA B51. The Canadian Registration Number (CRN) is:

0F09333.23 **Expiry Date: Feb. 12, 2029**

Please note that every fitting shall be constructed in strict accordance with the registered design.

Fitting registrations are required to be resubmitted for validation after ten (10) years from the registration date in accordance with CSA B51, Clause 4.2.1.

Should you require anything further, please do not hesitate to contact the Codes and Standards Compliance Office at your convenience.

Yours truly,



Williams Uju, P.Eng.
Codes and Standards Compliance

Remarks:

CRN registered under reciprocal agreement and conditional upon compliance with the notes on the original registration.