



Declaration of Conformity

Sierra Instruments, Inc.
5 Harris Court, Building L
Monterey, CA 93940

Sierra Instruments certifies that the products:

RedyCompact Meter GCM[‡] / Mass Flow Meter
RedyCompact Switch GCS[‡] / Mass Flow Meter
RedyCompact Regulator GCR[‡] / Mass Flow Controller
RedyCompact all-in GCA[‡] / Mass Flow Controller

[‡]All various versions.

Comply with the regulations of the following European Directives:

EMC-directive 2014/30/EU

Conformity is demonstrated by compliance with the following standards and normative documents:

IEC 61000-4-2 (2008-12 ed 2.0)
IEC 61000-4-3 (2010-04 ed 3.2)
IEC 61000-4-4 (2012-04 ed 3.0)
IEC 61000-4-5 (2014-05 ed 3.0)
IEC 61000-4-6 (2013-10 ed 4.0)
IEC 61326-1 (2012-07 ed 2.0)
CISPR11 (2010-05 ed 5.1)

Jerald W Mullison

Jerry Mullison

Director of Engineering

United States, June 2023