




Declaration of Conformity

Sierra Instruments, Inc.
5 Harris Court, Building L
Monterey, CA 93940

Sierra Instruments certifies that the products:

RedyIndustrial Meter GIM[‡] / Mass Flow Meter
RedyIndustrial Controller GIC[‡] / Mass Flow Controller



[‡]All various versions.

Comply with the regulations of the following European Directives:

EMC-directive 2014/30/EU
Ex-directive 2014/34/EU

Conformity is demonstrated by compliance with the following standards and normative documents:

EN55011 (2009) + A1 (2010)
EN61000-3-2 (2006) + A1 (2009) + A2 (2009)
EN61000-3-3 (2008)
EN61000-4-2 (2009)
EN61000-4-3 (2006) + A1 (2008) + A2 (2010)
EN61000-4-4 (2012)
EN61000-4-5 (2007)
EN61000-4-6 (2014)
EN61000-4-8 (2010)
EN61326-1 (2006)
NAMUR NE21 (2012)

Jerry Mullison

Jerry Mullison, Director of Engineering
United States, June 2023