



Declaration of Conformity

Sierra Instruments, Inc.
5 Harris Court, Building L
Monterey, CA 93940

Sierra Instruments certifies that the products:
RedySmart GSM[‡] / Mass Flow Meter
RedySmart Controller GSC[‡] / Mass Flow Controller
RedySmart-Mini Meter GMM[‡] / Mass Flow Meter
RedySmart-Mini Controller GMC[‡] / Mass Flow Controller
RedySmart Pressure Controller GSP[‡] / Pressure Controller
RedySmartm Back Pressure Controller GSB[‡] / Back Pressure Controller

[‡]All various versions.

Comply with the regulations of the following UK legislation:

S.I. 2019/1246
S.I. 2020/852
S.I. 2016/1091
S.I. 2017/1206
S.I. 2012/3032

Conformity is demonstrated by compliance with the following standards and normative documents:

IEC61000-4-2 (2008-12 ed2.0)
IEC61000-4-3 (2010-04 ed3.2)
IEC61000-4-4 (2012-04 ed3.0)
IEC61000-4-5 (2014-05 ed3.0)
IEC61000-4-6 (2013-10 ed4.0)
IEC61000-6-2 (2005-01 ed2.0)
IEC61326-1 (2012-07 ed2.0)
CISPR11 (2015-01 ed6.0)

Gerald W Mullison
Jerry Mullison, Director of Engineering
United States, June 2023